

Original Article

Development of A Loop-Mediated Isothermal Amplification (LAMP) Assay for Detection of Relapsing Fever *Borreliae*

Faezeh Houmansadr¹; *Mohammad Soleimani^{2,3}; *Saied Reza Naddaf⁴

¹Department of Cellular and Molecular Biology, Science and Research Branch, Islamic Azad University, Tehran, Iran

²Department of Microbiology, Faculty of Medicine, AJA University of Medical Sciences, Tehran, Iran

³Tasnim Biotechnology Research Center, Faculty of Medicine, AJA University of Medical Sciences, Tehran, Iran

⁴Department of Parasitology, Pasteur Institute of Iran, Tehran, Iran

(Received 18 Mar 2019; accepted 04 Mar 2020)

Abstract

Background: This study aimed to develop a loop-mediated isothermal amplification (LAMP) assay for the rapid detection of tick-borne relapsing fever in resource-limited areas.

Methods: A set of six primers were designed based on the conserved regions of the Glycerophosphodiester phosphodiesterase (*glpQ*) gene of *Borrelia* species. For sensitivity assay, serial dilutions of a recombinant plasmid containing a 219bp sequence of the *glpQ* were prepared and used as the template DNA. The LAMP reactions containing the six primers and the reagents required for amplification were incubated at 60–65 °C for 60min in a Loopamp real-time turbidimeter. For the specificity test, DNA from 14 other bacteria were included in the assays, and double-distilled water was used as the negative control. Also, DNA from dried blood spots (DBSs) of spirochetemic mice, and blood samples from relapsing fever-suspected patients were examined by the LAMP along a *Borrelia*-specific nested PCR that targets the *rrs-rrl*-IGS region.

Results: The LAMP detected as low as 90*glpQ* copies in reactions. The primers reacted with DNA from DBS of spirochetemic mice showing spirochete concentrations of \leq one per a 1000X microscopic field. In clinical samples, the LAMP assay showed a higher sensitivity compared to nested-PCR. The LAMP specificity was 100%, as the primers did not react with other bacteria DNA.

Conclusion: The high sensitivity and specificity of the test, along with the simplicity of the DNA extraction procedure, make the LAMP a reliable and adaptable tool for the diagnosis of tick-borne relapsing fever in rural endemic areas.

Keywords: Relapsing fever; Loop-Mediated Isothermal Amplification (LAMP); Iran

Introduction

The genus *Borrelia* comprises two distinct groups of spirochetes with the difference in diseases they cause. One group includes the causative agents of Lyme disease and the other the relapsing fever borreliae (RFB). Currently, there are 22 confirmed RFB, and six other taxa have been proposed (1). Except for the louse adapted *Borrelia recurrentis*, the majority of species pathogenic to humans are transmitted by the soft ticks of the genus *Ornithodoros*. A few species, such as *Borrelia miyamotoi* and *Borrelia lonestari*, are vectored by

hard ticks and yet share genetic similarities with the RFB (2). Soft tick-borne relapsing fever (STBR) is endemic to Iran (3-5). Until now, despite the improvement of housing and the removal of the disease from mandatory reporting to the Ministry of Health and Medical Education (MHME), no year has passed without reports of human infections. In Iran, four RFB, including *Borrelia persica*, *Borrelia microti*, *Borrelia latyschewii* and *Borrelia baltazardi* have been described (5). *Borrelia persica* is the primary cause of the disease especially in the

*Corresponding authors: Dr Mohammad Soleimani, E-mail: soleimanidor@yahoo.com, Dr Saied Reza Naddaf, E-mail: saiedrezanaddaf@gmail.com

west and northwest of the country (5-7), while in the south, epidemiological data and molecular approved human infections indicated *B. microti* and other *B. microti*-like borreliae as the other cause of relapsing fever (3, 8, 9).

In Iran, until recently, confirmation of relapsing fever merely relied on observation of the spirochetes in peripheral blood of febrile patients using darkfield microscopy or Giemsa-stained blood smears. The disease is easily diagnosed by microscopy during fever peaks with a massive spirochetemia. However, between the peaks and in milder diseases, the bacteria are scanty and are hard to identify in blood smears. In Iran, PCR assays using various molecular markers like *flaB*, *glpQ* and *rrs* have successfully detected the *Borrelia* spp in relapsing fever patients (3, 8) and animals (10). The PCR assays exhibit high sensitivities but are not commonly affordable in resource-limited laboratories of rural areas, where most of the relapsing fever infections occur. Hence, we prompted to develop an alternative DNA amplification assay of lower cost for the detection of the disease in these areas. Loop-mediated isothermal amplification (LAMP) has proved as a robust, cheap, highly sensitive/specific tool for the detection of various pathogenic agents including viruses, bacteria, fungi, and parasites (11-14). This assay employs a DNA polymerase and a set of 4–6 primers that operate in isothermal conditions forming loop structures that ultimately precipitate in the reaction mixture (13). This assay has also shown to be less prone to inhibition from DNA preparations and allows adaptability to field conditions (15, 16).

This study aimed to develop a (LAMP) assay for the rapid detection of relapsing fever borreliae based on the glycerophosphodiester phosphodiesterase (*glpQ*) gene, a sequence conserved among all relapsing fever borreliae, but absent from Lyme disease spirochetes (17).

Materials and Methods

Bacteria species and DNA extraction

We used *B. microti* strain IR-1, which was maintained via continual passages in NMRI mice for more than 15 years in the Parasitology Department of Pasteur Institute of Iran. Blood samples were obtained from *B. microti*-infected mice when spirochetes reached 1.4×10^6 /ml of blood. DNA extraction from 1ml of blood was performed using a Genomic DNA Purification Kit (Promega, Madison, USA) as described by the manufacturer.

Primer design

Initially, *glpQ* sequences of 10 relapsing fever *Borrelia* (*B. microti*, Acc. No. JF825473; *B. microti*, Acc. No. EU914144; *B. recurrentis*, Acc. No. KJ003842, *Borrelia* sp., Acc. No. KX683865; *Borrelia duttonii*, Acc. No. DQ346785; *Borrelia crocidurae*, Acc. No. CP004267; *Borrelia hispanica*; Acc. No. GU357573; *B. persica*, Acc. No. EU914143; *B. recurrentis*, Acc. No. AF247152; *Borrelia duttonii*, Acc. No. GU357577) were obtained from the GenBank database and aligned by using CLC Sequence Viewer 7 (CLC bio, Aarhus, Denmark). A set of six primers including two loop primers were designed based on the conserved regions of the *glpQ* sequence, corresponding to the nucleotides 261015–261682 of *B. duttonii* strain Ly (Fig. 1), by the online software program, Primer Explorer V4 (Eiken Chemical Co., Tokyo, Japan; <http://primerexplorer.jp/e/>). The theoretical specificity of the designed primers was confirmed by in silico analysis using BLAST and Primer-BLAST software available in NCBI (<http://www.ncbi.nlm.nih.gov/>). The primers were synthesized by a commercial company (Generay Biotechnology, Shanghai, China). The main primers (*glpQ*-F3 and *glpQ*-B3) and (*glpQ*-FIP and *glpQ*-BIP) target fragments of 219bp and 161bp size, respectively, and the loop primer (*glpQ*-LF and *glpQ*-LB) produce

amplicons of various size with a ladder-like pattern (Table 1).

Glycerophosphodiester Phosphodiesterase Gene (*glpQ*) cloning

A 219bp fragment of the *glpQ* gene was amplified by the primers *glpQ*-F3 and *glpQ*-B3 (Table 1). The reaction contained 3mM MgSO₄ (Biobasic, Toronto, Canada), 1.6mM dNTPs (Kawsar Biotech Co, Tehran, Iran), 1μl 10X buffer [100mM KCl, 100mM (NH₄)₂SO₄, 200mM Tris HCl (pH 8.75)], 1% Triton X-100, 1mg/ml BSA (Biobasic, Toronto, Canada), 1U Taq DNA Polymerase (Biobasic, Toronto, Canada), 0.4μM of each primer, 1μl template DNA, and double-distilled water (DDW) to the 25μl final volume. The amplification was programmed in a thermal cycler (Eppendorf, Hamburg, Germany), for an initial denaturation at 94 °C for 4min followed by 35 cycles of 94 °C for 45sec, 45 °C for 45sec, and 72 °C for 30sec, and a final extension at 72 °C for 10min. The PCR products were run on a 2% agarose gel (Min Run Gel Electrophoresis System; Bio-Equip co, Shanghai, China), stained with ethidium bromide (CinnaGen, Alborz, Iran), and visualized under UV in Gel Documentation system (E-BOX VILBER, Marne-la-Vallée, France). The PCR product was purified using a PCR Purification Kit (Bioneer, Daejeon, South Korea), cloned into a TA vector (InstaClone™ PCR Cloning Kit, Thermo Scientific, MA, United States), and transformed in *Escherichia coli* Top10F'. The bacteria were incubated at 37 °C for 24h on Luria-Bertani medium (Merck, KGaA, Darmstadt, Germany) containing 24mg/ml IPTG (isopropyl-beta-D-thiogalactopyranoside) (Fermentas, Ontario, Canada), 20mg/ml X-gal (5-bromo-4-chloro-3-indolyl beta d-galactoside) (Fermentas, Ontario, Canada), 10mg/ml tetracycline (Razak, Alborz, Iran) and 50mg/ml ampicillin (Cosar, Tehran, Iran). The recombinant bacteria were identified by blue/white screening, with the white colonies representing recombinant ones. One white colony was added to Luria-Bertani

broth medium containing 10mg/ml tetracycline and 50mg/ml ampicillin followed by incubation at 37 °C for 16h while shaking at 180 RPM. Plasmid purification was performed by the AccuPrep Plasmid Mini Extraction kit (Bioneer, Daejeon, South Korea) and the presence of the *glpQ* gene in the recombinant plasmids was confirmed by PCR amplification with the primers *glpQ*-F3 and *glpQ*-B3. The recombinant plasmid was named pTZ57R/T-*glpQ*.

Sensitivity assay

A serial 10-fold dilution of the recombinant plasmid ranging from 9×10^8 to 9×10^{-1} copy numbers per microliter, equivalent to $\approx 32 \text{ ng}$ to $\approx 32 \times 10^4 \text{ fg}/\mu\text{l}$ of DNA was prepared and used in assays.

Specificity assay

For specificity assays, we used DNA of 14 other bacteria including *Shigella sonnei* ATCC 9290, *Klebsiella pneumoniae* ATCC 7881, *Bacillus subtilis* ATCC 6051, *Staphylococcus aureus* ATCC 25923, *Enterococcus faecalis* ATCC 29212, *Enteropathogenic Escherichia coli* (EPEC) ATCC 43887, *Yersinia enterocolitica* ATCC 23715, *Pseudomonas aeruginosa* ATCC 27853, clinical specimens of *Escherichia coli*, *Salmonella typhi*, *Acinetobacter baumannii*, *Citrobacter* sp, *Enterobacter* sp, and *Leptospira interrogans* in all assays. All the bacteria, including *L. interrogans*, the closely related bacteria to *Borrelia* species contained *glpQ* sequence.

Loop Mediated Isothermal Amplification (LAMP) assay

The LAMP reactions contained 40pM of the inner (*glpQ*-FIP and *glpQ*-BIP) and 10pM of outer (*glpQ*-F3 and *glpQ*-B3) and loop primers (*glpQ*-LF and *glpQ*-LB) (Table 1), 11.2 mM dNTPs (Kawsar Biotech Co, Tehran, Iran), 0.8M betaine (Sigma Aldrich, Taufkirchen, Germany), 20mM Tris-HCl, 10mM KCl, 10mM (NH₂)SO₄, 0.05% Triton X-100 (pH 8.8) (Biolabs, New England, UK), 8mM MgSO₄ (Bi-

obasic, Toronto, Canada), 0.1% Tween 20 (Acros Organics, Vernon, USA), 8U of Bst DNA polymerase, large fragment (Biolabs, New England, UK), 1µl of serial dilutions of the recombinant plasmid, and DDW to the 25 µl final volume. The reactions were incubated at temperatures ranging from 60–65 °C for 60min in a Loopamp real-time turbidimeter (LA-320C; Teramecs, Kyoto, Japan), followed by heating at 80 °C for 5min for enzyme inactivation. In all assays, DNAs from 14 other bacteria were included and DDW was used as a negative control.

During the optimization of the assay, some amplifications were performed without the loop primers. Also, in some reactions, we added 1µl of Fluorescent detection reagent containing Calcein (Eiken Chemical co., Tokyo, Japan), an indicator of DNA amplification.

Detection of LAMP products

The amplification in LAMP reactions was examined by 1) naked eye observation of white turbidity resulting from the accumulation of magnesium pyrophosphate, a by-product of the reactions, 2) a Loopamp real-time turbidimeter that records the optical density of reactions every 6sec at 650nm (the reactions were considered positive when the turbidity reached ≥ 0.1 within 60min), 3) the color change from orange to green, as an indication of DNA amplification, and 4) gel electrophoresis of amplicons on 2% agarose gels.

A LAMP product resulting from the 9×10^8 dilution was purified using PCR Purification Kit (Bioneer, Daejeon, Korea) and sequenced in both directions by (ABI 3730xl/Bioneer 3730xl, Daejeon, Republic of Korea).

Preparation of DBSs from *Borrelia*-infected mice

Amounts of 200µl blood from *B. microti*-infected mice with various degrees of the spirochetemia (225 ± 61.34 , 71 ± 36.71 , 3.3 ± 1.63 , 1.1 ± 1.44 , and 0.8 ± 0.78 spirochetes per microscopic field) were dotted on 30 DNA banking

cards (DBC) (Kawsar Biotech Co, Tehran, Iran). Blood from non-infected mice was used as controls on DBCs. The blood spots were allowed to dry at room temperature, and the DBCs were kept in the same condition until used.

Loop Mediated Isothermal Amplification (LAMP) and PCR with DBS

Circles of 2mm from DBSs were cut and washed three times with DNA extraction buffer (provided by the manufacturer), followed by DDW. The circles were allowed to dry at room temperature and then used in LAMP and PCR assays as described above except that instead of template DNA, DBS circles were included in the reactions.

The sensitivity and specificity of the LAMP assay were calculated in comparison with microscopy as the gold standard assay using MedCalc (2018 MedCalc Software bvba) software available online (https://www.medcalc.org/calc/diagnostic_test.php).

Loop Mediated Isothermal Amplification (LAMP) and nested-PCR amplification of clinical specimens

DNA was extracted from 39 sera of febrile patients residing in the relapsing fever endemic areas in the south and west of Iran using a commercial DNA extraction kit as described above. The DNA samples were examined by the LAMP, and a *Borrelia*-specific nested PCR that amplifies the *rrs-rrl*-IGS region using the outer primers F, 5'-GTATGTTAGTGAGGGGGGTG-3' and R, 5'-GGATCATAGCTCAGGTGGTTAG-3' and inner nested primers F, 5'-AGGGGGGTGAAGTCGTAACAAG-3' and R, 5'-GTCTGATAAACC TGAGGTCGGA-3' (18). This PCR has exhibited high sensitivity in detecting relapsing fever borreliae (3, 8, 19, 20). In all assays, for specificity test, DNAs from other bacteria species were included and DDW was used as a negative control. The LAMP reactions were checked for DNA amplification as described above, and the

PCR products were resolved on 2% agarose gels, visualized under UV and photographed.

Results

Glycerophosphodiester Phosphodiesterase Gene (*glpQ*) cloning

Amplification of the *glpQ* sequence from the recombinant plasmid (pTZ57R/T-*glpQ*) with the primers *glpQ*-F3 and *glpQ*-B3 yielded the expected 219bp indicating the insertion of this sequence in the plasmid.

Loop Mediated Isothermal Amplification (LAMP) assay with recombinant pTZ57R/T-*glpQ*

We observed amplification of *glpQ* sequence in the LAMP reactions by naked eye observation of turbidity and the Loopamp real-time turbidimeter down to 9×10^3 copies of recombinant plasmid equivalent to 0.32fg DNA (Fig. 2A). However, gel electrophoresis revealed amplifications in two lower dilutions of 9×10^2 and 9×10^1 , equivalent to 3.2fg and 0.32fg DNA, respectively (Fig. 2B). Our LAMP assay was 100% specific as the primers reacted with none of the other 14 bacterial DNA, including *L. interrogans*. In the reactions containing calcein and the *Borrelia* DNA, the color turned from orange to green.

The 137bp sequence resulted following the sequencing of the LAMP amplicon matched with the *glpQ* sequence of *B. microti* strain IR-1 (acc. No. JF825473) corresponding to nucleotides 261158–261295 of the whole genome

sequence of *B. duttonii* Ly (acc. No. CP000796). The results were the same within the temperature range of 60–65 °C. In the reactions, in absence of loop primers, the optimum time for isothermal amplification was 60min, whereas, in those with the loop primers, the incubation time reduced to 45min.

Loop Mediated Isothermal Amplification (LAMP) and PCR with DBSs

Of the 30 *B. microti*-positive DBS, all (100%) showed turbidity with the naked eye indicating amplification, while the *glpQ*-PCR (with the primers *glpQ*-F3 and *glpQ*-B3) yielded the expected 219bp band in 20 positive DBSs (66.67%). The negative PCR reactions belonged to the mice with the lowest level of spirochetemia [1.1 ± 1.44 (n= 4) and 0.8 ± 0.78 (n= 6) spirochete per microscopic field]. Neither LAMP nor PCR amplification was observed with the negative controls (DDW) or reactions containing other bacteria DNA as the template.

By considering microscopy as the gold standard, the sensitivity of *glpQ*-LAMP and *glpQ*-PCR were 100% and 66.67%, respectively. The specificity was 100% for both assays.

Loop Mediated Isothermal Amplification (LAMP) and PCR with clinical samples

Of 39 clinical samples, 11 became turbid in the LAMP assay indicating amplification of *glpQ* gene, whereas PCR amplification of IGS sequence only yielded the expected 540bp band in three specimens (Fig. 3).

Table 1. The primers designed and used for amplification of *glpQ* gene by LAMP assay

Primers	Name	Sequences (5' to 3')
<i>glpQ</i> -F3	Forward outer primer	AATGCACGATCCTGAAC
<i>glpQ</i> -B3	Backward outer primer	TCTTCTTCTAGGGTTGGAATT
<i>glpQ</i> -FIP	Forward inner primer	TGCTAATGTGAAATCGACGGAATAA- CAACAACAAATGTTGCAAAGC
<i>glpQ</i> -BIP	Backward inner primer	AATCACTAAGCCTTAGCGAAAGAT- TGTTGCAGGAAAACGGTTA
<i>glpQ</i> -LF	Forward loop primer	TCTCTAGCTCTTCCTGGAAACA
<i>glpQ</i> -LB	Backward loop primer	CCTGAAACACAACAACCAATATACC

ters (5), and overlooking the spirochetes in the blood of some febrile patients did not question the identity of the causative agent. However, with the decline of the disease in the endemic regions and reports of sporadic cases from other areas application of more sensitive approaches for the identification RFB became necessary. In Iran, over the past decade, PCR assays by based on various molecular markers, such as *glpQ*, *rrs*, and *flaB*, were developed for the identification of *Borrelia* infection in *Ornithodoros* ticks (21, 22) or characterization of the tick-originated relapsing fever borreliae (9, 23), but rarely clinical samples were included. Lately, qPCR and conventional PCRs identified a *B. microti*-like strain, presumably, an ecotype of African *B. duttonii*, in relapsing fever patients from southern Iran (3, 8).

In our previous work, using *B. persica*-spiked blood samples, we consistently observed bacteria by microscopy in the blood samples with densities ≥ 800 –1000 spirochetes/ μl (24). Herein, with the DBSs prepared with the spirochetemic murine blood, the LAMP could detect borrelial DNA in blood specimens showing less than one spirochetes per 1000X microscopic field. Our LAMP assay also showed a higher sensitivity in comparison with a nested-PCR amplification of the IGS in the diagnosis of clinical samples. The specificity of the LAMP assay for the diagnosis of *B. microti* DNA was 100% as the designed primers exhibited no cross-reaction with DNA from the other 14 bacteria used in this study. Our LAMP method requires to be further tested with blood samples from RF patients infected with other *Borrelia* species to ensure the sensitivity and specificity of the assay.

LAMP method has also shown promise for the detection of borreliae DNA and other pathogens in the tick vectors. In China, the LAMP could detect *B. burgdorferi* s. l. in ticks, with a higher sensitivity than a conventional PCR (25). Also, in recognizing spotted fever group

rickettsia, the LAMP appeared ten times more sensitive than an end-point PCR targeting the same gene fragment (26).

Conclusion

Such a reliable sensitivity and specificity, as well as the simplicity of the DNA extraction procedure, make the LAMP a suitable and adaptable assay for field studies in relapsing fever endemic areas. In resource-limited rural health centers, LAMP can readily check field-collected blood samples, preferably on DBS, for relapsing fever borreliae with the least equipment accessible, i.e., a heat block. Having access to a Loopamp real-time turbidimeter is preferable. However, LAMP is a qualitative assay, and the result can also be monitored by color change with calcein in the fluorescent detection reagent or by the turbidity of magnesium pyrophosphate in the reaction tubes. Hence, a heating system that provides a temperature within the range of 60–65 °C would suffice.

Acknowledgements

We thank the Tasnim Biotechnology of Research Center (TBRC), AJA University of Medical Sciences, for their technical support. The authors declare that there is no conflict of interest.

References

1. Talagrand-Reboul E, Boyer PH, Bergstrom S, Vial L, Boulanger N (2018) Relapsing Fevers: Neglected Tick-Borne Diseases. *Front Cell Infect Microbiol.* 8: 98.
2. Cutler SJ (2015) Relapsing Fever Borreliae: A Global Review. *Clin Lab Med.* 35 (4): 875–865.
3. Naddaf SR, Ghazinehad B, Sedaghat MM, Asl HM, Cutler SJ (2015) Tickborne re-

- lapsing fever in southern Iran, 2011–2013. *Emerg Infect Dis.* 21(6): 1078–1080.
4. Masoumi Asl H, Goya MM, Vatandoost H, Zahraei SM, Mafi M, Asmar M, Piazak N, Aghighi Z (2009) The epidemiology of tick-borne relapsing fever in Iran during 1997–2006. *Travel Med Infect Dis.* 7(3): 160–164.
 5. Karimi U (1981) Relapsing Fever and its Epidemiology. Pasteur Institute of Iran Press, Tehran (In Persian).
 6. Karimi Y, Hovind-Hougen K, Birch-Andersen A, Asmar M (1979) *Borrelia persica* and *B. baltazardi* sp. nov.: experimental pathogenicity for some animals and comparison of the ultrastructure. *Ann Microbiol.* 130B(2): 157–168.
 7. Karimi Y, Hannoun C, Ardouin P, Ameli M (1976) Sur le purpura hé-morragique observe 'dans l'Azerbaïdjan-Est de l'Iran. *Me'd Mal Infec.* 6: 399–404.
 8. Naddaf SR, Ghazinezhad B, Kazemirad E, Cutler SJ (2017) Relapsing fever causative agent in Southern Iran is a closely related species to East African borreliae. *Ticks Tick-Borne Dis.* 8(6): 882–886.
 9. Naddaf SR, Ghazinezhad B, Bahramali G, Cutler SJ (2012) Phylogenetic analysis of the spirochete *Borrelia microti*, a potential agent of relapsing fever in Iran. *J Clin Microbiol.* 50(9): 2873–2876.
 10. Shirani D, Rakhshanpoor A, Cutler SJ, Ghazinezhad B, Naddaf SR (2016) A case of canine borreliosis in Iran caused by *Borrelia persica* *Tick Tick-Borne Dis.* 7(3): 424–426.
 11. Han F, Ge B (2010) Quantitative detection of *Vibrio vulnificus* in raw oysters by real-time loop-mediated isothermal amplification. *Int J Food Microbiol.* 142(1): 60–66.
 12. McKenna JP, Fairley DJ, Shields MD, Cosby SL, Wyatt DE, McCaughey C, Coyle PV (2011) Development and clinical validation of a loop-mediated isothermal amplification method for the rapid detection of *Neisseria meningitidis*. *Diag Microbiol Infect Dis.* 69(2): 137–144.
 13. Notomi T, Okayama H, Masubuchi H, Yonekawa T, Watanabe K, Amino N, Hase T (2000) Loop-mediated isothermal amplification of DNA. *Nucleic Acids Res.* 28(12): 1–7.
 14. Mansour SM, Ali H, Chase CC, Cepica A (2015) Loop-mediated isothermal amplification for diagnosis of 18 World Organization for Animal Health (OIE) notifiable viral diseases of ruminants, swine and poultry. *Anim Health Res Rev.* 16(2): 89–106.
 15. Caipang CMA, Haraguchi I, Ohira T, Hirono I, Aoki T (2004) Rapid detection of a fish iridovirus using loop-mediated isothermal amplification (LAMP). *J Virol Methods.* 121(2): 155–161.
 16. Mori Y, Notomi T (2009) Loop-mediated isothermal amplification (LAMP): a rapid, accurate, and cost-effective diagnostic method for infectious diseases. *J Infect Chemother.* 15(2): 62–69.
 17. Schwan TG, Schrumph ME, Hinnebusch BJ, Anderson DE, Jr, Konkel ME (1996) GlpQ: an antigen for serological discrimination between relapsing fever and Lyme borreliosis. *J Clin Microbiol.* 34(10): 2483–2492.
 18. Bunikis J, Garpmo U, Tsao J, Berglund J, Fish D, Barbour AG (2004) Sequence typing reveals extensive strain diversity of the Lyme borreliosis agents *Borrelia burgdorferi* in North America and *Borrelia afzelii* in Europe. *Microbiol.* 150(6): 1741–1755.
 19. Scott JC, Wright DJM, Cutler JC (2005) Typing African relapsing fever spirochetes. *Emerg Infect Dis.* 11(11): 1722–1729.
 20. Cutler SJ, Bonilla EM, Singh RJ (2010) Population structure of East African relapsing fever *Borrelia* spp. *Emerg Infect*

- Dis. 16(7): 1076–1080.
21. Rafinejad J, Choubdar N, Oshaghi M, Piazak N, Satvat T, Mohtarami F, Barmaki A (2011) Detection of *Borrelia persica* Infection in *Ornithodoros tholozani* Using PCR Targeting rrs Gene and Xenodiagnosis. Iran J Public Health. 40(4): 138–145.
 22. Barmaki A, Rafinejad J, Vatandoost H, Telmadarraiy Z, Mohtarami F, Leghaei Sh, Oshaghi M (2010) Study on Presence of *Borrelia persica* in Soft Ticks in Western Iran. Iran J Arthropod Borne Dis. 4 (2): 19–25.
 23. Oshaghi MA, Rafinejad J, Choubdar N, Piazak N, Vatandoost H, Telmadarraiy Z, Mohtarami F, Ravasan NM (2011) Discrimination of relapsing fever *Borrelia persica* and *Borrelia microtti* by diagnostic species-specific primers and polymerase chain reaction-restriction fragment length polymorphism. Vector Borne Zoonotic Dis. 11 (3): 201–207.
 24. Naddaf SR, Kishdehi M, Siavashi M (2011) Comparison of PCR-Based Diagnosis with Centrifuged-Based Enrichment Method for Detection of *Borrelia persica* in Animal Blood Samples. Iran J Arthropod Borne Dis. 5(1): 7–12.
 25. Yang J, Guan G, Niu Q, Liu Z, Li Y, Liu J, Ma M, Ren Q, Liu A, Yin H (2013) Development and application of a loop-mediated isothermal amplification assay for rapid detection of *Borrelia burgdorferi* s. l. in ticks. Transbound Emerg Dis. 60(3): 238–244.
 26. Noden BH, Martin J, Carrillo Y, Talley JL, Ochoa-Corona FM (2018) Development of a loop-mediated isothermal amplification (LAMP) assay for rapid screening of ticks and fleas for spotted fever group rickettsia. PLoS One. 13(2): e0192331.